



Industrie Service

The Testing and Calibration Laboratory
of TÜV SÜD Industrie Service GmbH

confirm to the company

Disruptive Technologies Research AS

Ytrebygdsvegen 215

NO-Blomsterdalen, 5258

because of the positive results of the tested temperature recorder,

Cloud Connector EU 4G with wireless sensors

according to the test-report No. **1132 TS** from the 2020-01-07, that the temperature recorder fulfils the requirements of the standard DIN EN 12830:2018-10.

accuracy class	0,5	1
measurement range	0 °C to +55 °C	-25 °C to +55 °C
application range	storage	
software verification level	III	

The manufacturer is allowed to mark the temperature recorder with the following test-report number:

ATP-MUC 1132 TS

This certificate
is valid to: **2025-01-06**

(As long as the normative requirements, the declared fields of application, the product and the production conditions are not significantly changed)

Munich, the 2020-01-07

Center of Competence
Refrigeration and air-conditioning
Head of test laboratory

Expert

Andreas Klotz



Thorsten Bennert